



OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R60/2000-NL1-11.35
Project number 11200427
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. NMI-11200427-01 dated 21 February 2012 that includes 27 pages;
- No. NMI-11200427-02 dated 21 February 2012 that includes 24 pages.

Characteristics of the Load Cell:

Fraction P_i	0,7	
Maximum capacity (E_{max})	100 kg up to and including 2500 kg	
Humidity Class	CH	
Temperature range	-10 °C / +40 °C	
Accuracy Class	C	
Maximum number of load cell intervals (n)	6000	5000
Ratio of minimum LC Verification interval $Y = E_{max} / V_{min}$	15000	15000
Ratio of minimum dead load output return $Z = E_{max} / (2 * DR)$	7000	5000